

Display Report

Analysis Info

Analysis Name D:\Data\test\0424\leigongtengneizhijia_73_01_1124.d
Method 1124.m
Sample Name leigongtengneizhijia
Comment Agilent ZORBAX SB-C18 5um 4.6*250mm
ACN: Water=90:10

Operator
Instrument braker
amaZon SL

Acquisition Parameter

Ion Source Type	ESI	Ion Polarity	Positive	Alternating Ion Polarity	on
Mass Range Mode	UltraScan	Scan Begin	70 m/z	Scan End	1000 m/z
Capillary Exit	140.0 V	n/a	n/a	Trap Drive	50.6
Accumulation Time	8629 μ s	Averages	5 Spectra	Auto MS/MS	off

